

EU Declaration of Conformity

Manufacturer Name: Kingstar Medical (Xianning) Co., Ltd.

Manufacturer Address: No. 79 Yong'andong Road, Xian'an District 437100, Xianning City, Hubei Province, the People's Republic of China

SRN of the Manufacturer: CN-MF-000006015

Location of Manufacturer: Xianning City, Hubei Province, China.

Authorized Representative: Shanghai International Holding Corp. GmbH (Europe) [DE]

SRN of the Authorized Representative: DE-AR-000000001

Address of their Registered Place of Business: Eiffestraße 80, 20537 Hamburg, Germany

Location be established: Germany

Basic UDI-DI: 6971872201171000N3

Trade Name: Nonsterile Self-adhesive Bandages

Name of the Device: Nonsterile Self-adhesive Bandages

EMDN Code: M03040202, Elastic Support Bandages, Cohesive (Self-Adhesive)

GMDN Code: 58964, Non-adhesive device retention bandage

Intended Purpose: It is primarily used to provide a restraint force to the dressing or limb to function as a bandaging, securement.

Risk Class of the Device: Class I, based on Rule 1 of ANNEX VIII of Regulation (EU) 2017/745.

- All non-invasive devices are classified as class I.

The conformity assessment procedure performed: According to Article 19 of the Regulation (EU) 2017/745, draw up this EU declaration of conformity which contain the information set out in Annex II and III of the Regulation (EU) 2017/745.

CS used or Standard applied: Please find in Annex II.

Identification of the device: Please find in Annex I.

Declaration: This declaration of conformity is issued under the sole responsibility of Kingstar Medical (Xianning) Co., Ltd. We hereby declare that the medical device specified above meet the provision of the Regulation (EU) 2017/745 for medical device. This declaration is supported by the quality system approval to ISO 13485 by TÜV SÜD Product Service GmbH.

All supporting documentation is retained at the premises of the manufacturer.

Signed for and on behalf of:

Place of Issue: Xianning City, Hubei Province, China.

Date of Issue: 2025-06-09

Print Name: Fan Rong

Function: Management Representative

Signature: 

Annex I --- Identification of the Device Covered by the EU Declaration of Conformity

1. Identification of the Device

Table --- Identification of the Device

No.	WERO REF No.	Kingstar REF No.	Kingstar Art. Nr.	Product Name	Specification	Packaging Configuration
1	201029	C020003	5142370010	Nonsterile Self-adhesive Bandages	3cm×5.5m	1 roll/foil bag, 250 rolls/carton
2	201031	C020003	5142370020	Nonsterile Self-adhesive Bandages	3cm×5.5m	1 roll/foil bag, 12 bags/box, 50 boxes/carton

2. Photograph of Nonsterile Self-adhesive Bandages



Photo 1 --- Self-adhesive Bandages



Photo 2 --- Self-adhesive Bandages in nonsterile packaging

Annex II --- European Harmonization and International Standard list

No.	Standards	Content
1	EN ISO 13485:2016/A11: 2021	Quality management systems — Requirements for regulatory purposes (ISO 13485:2016)
2	ISO 15223-1: 2021/Amd1: 2025	Medical devices - Symbols to be used with information to be supplied by the manufacturer- Part1: General requirements - Amendment 1: Addition of defined term for authorized representative and modified EC REP symbol to not be country or region specific
3	EN ISO 20417: 2021	Medical devices- Information to be supplied by the manufacturer (ISO 20417:2021)
4	EN ISO 14971:2019/A11: 2021	Medical devices-Application of risk management to medical devices (ISO 14971:2019)
5	IEC 62366-1:2015/A1: 2020	Medical devices-Application of usability engineering to medical devices - Amendment 1
6	MEDDEV 2.7.1 rev 4	CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES UNDER DIRECTIVES 93/42/EEC and 901385/EEC
7	ISO 2859-1:1999/A1: 2011	Sampling procedures for inspection by attributes--Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection - Amendment 1
8	ISTA 2A: 2011	Partial simulation performance test procedure
9	MEDDEV 2.12-1: 2013	Guidelines on a medical devices vigilance system
10	ISO/TR 20416: 2020	Medical devices - Post-market surveillance for manufacturers
11	EN ISO 14644-1: 2015	Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration (ISO 14644-1:2015)
12	EN ISO 14644-2: 2015	Cleanrooms and associated controlled environments — Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration (ISO 14644-2:2015)
13	ISO 14644-3:2019/Amd 1: 2020	Cleanrooms and associated controlled environments — Part 3: Test methods (ISO 14644-3:2019)
14	EN ISO 10993-1: 2020	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2018)
15	EN ISO 10993-5: 2009	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)
16	EN ISO 10993-10: 2023	Biological evaluation of medical devices — Part 10: Tests for skin sensitization (ISO 10993-10:2021)

17	ISO 10993-18:2020/A1: 2023	Biological evaluation of medical devices - Part 18: Chemical characterization of medical device materials within a risk management process - Amendment 1: Determination of the uncertainty factor (ISO 10993-18:2020/Amd 1:2022)
18	EN ISO 10993-23: 2021	Biological evaluation of medical devices - Part 23: Tests for irritation (ISO 10993-23:2021)
19	REACH Regulation 1907/2006	Registration, Evaluation, Authorization and Restriction of Chemicals
20	CLP Regulation 1272/2008	Classification, Labelling and Packaging of Substances and Mixtures, Amending and Repealing Directives 67548/EEC and 1999/45/EC, and amending Regulation (EC) No 190712006